



广微测
Gmicro Testing

广东省微生物分析检测中心

GUANGDONG DETECTION CENTER OF MICROBIOLOGY

分析检测报告

REPORT FOR ANALYSIS

报告编号

Report №.

2021FM02724R03

样品名称

Name of Sample

中拓鼎力®CHPD-Y-F100 型空气消毒机

委托单位

Applicant

青岛陆海利达交通装备有限公司

检测类型

Test Type

委托检测

单位地址: 广州市先烈中路 100 号大院 66 号楼

Address:

Building 66, No.100 Central Xian Le Road, Guangzhou, China

邮政编码: 510070

Postcode:

电话号码: (020)87137666

Tel:

传真号码: (020)87137668

Fax:

网 址: www.gddcm.com

Website:



广微测
Gmicro Testing

广东省微生物分析检测中心

GUANGDONG DETECTION CENTER OF MICROBIOLOGY

分析检测报告

REPORT FOR ANALYSIS



报告编号 (Report No.) 2021FM02724R03 校验码 (Verification Code): 53819206

样品名称 Name of Sample	中拓鼎力®CHPD-Y-F100 型空气 消毒机	检测类型 Test Type	委托检测
委托单位 Applicant	青岛陆海利达交通装备有限公司	地 址 Address	青岛市城阳区棘洪滩街道后海西 村委西 1000 米
样品来源 Sample Source	委托方送检	样品数量 Sample Quantity	1 台
样品规格和批号 Spec and Lot No of Sample	CHPD-Y-F100	样品状态和特性 State and Characteristic	机器
接样日期 Sample Received Date	2021-02-04	检测完成日期 Completion Date	2021-03-10
检测依据和方法 Test Standard and Method	参照《消毒技术规范》2002 年版-2.1.3		
检测项目 Item Tested	空气除病毒效果鉴定试验		
检测结论 Test Conclusion	该样品所检项目的实测数据见本检测报告续页。 签发日期: 2021-03-24 Issue Date: 2021-03-24 机构盖章 Official Seal 检验检测专用章		
备注 Remarks			

制 表:
Editor

印婉仪

审 核:
Verifier

孙延丽

批 准:
Approver

孙延丽



广东省微生物分析检测中心

GUANGDONG DETECTION CENTER OF MICROBIOLOGY

分 析 检 测 结 果

ANALYSIS AND TEST RESULT

报告编号 (Report No.): 2021FM02724R03

实验病毒及宿主	作用时间	试验序号	空气病毒含量 (TCID ₅₀ /m ³)	杀灭率 (%)
甲型流感病毒 H1N1 (A/PR/8/34) 宿主名称: MDCK 细胞	0 (CK)	1	4.07×10 ⁶	
		2	5.13×10 ⁶	
		3	6.76×10 ⁶	
	1h	1	5.13×10 ³	99.83
		2	8.71×10 ³	99.76
		3	1.62×10 ⁴	99.68
甲型流感病毒 H3N2 宿主名称: MDCK 细胞	0 (CK)	1	5.13×10 ⁶	
		2	4.07×10 ⁶	
		3	5.13×10 ⁶	
	1h	1	1.38×10 ⁴	99.57
		2	8.71×10 ³	99.67
		3	8.13×10 ³	99.79
肠道病毒 71 型 宿主名称: Vero 细胞	0 (CK)	1	5.13×10 ⁶	
		2	6.76×10 ⁶	
		3	5.13×10 ⁶	
	1h	1	3.24×10 ⁴	99.21
		2	3.89×10 ⁴	99.24
		3	3.24×10 ⁴	99.17
注：杀灭率试验结果已消除微生物在空气中自然消亡因素的影响。 (以下空白)				
备注 Remarks	样品在 3m ³ 的试验舱内开启最大风速进行实验。			



报告编号 (Report №.): 2021FM02724R03

注意事项 Notice Items

1. 检测报告无本单位检验检测专用章、骑缝章无效。

The Test report is invalid if not affixed with Authorized Stamp of Test and Paging Seal.

2. 检测报告无审核人、批准人签字无效。

The Test report is invalid without signature of verifier and approver.

3. 检测报告涂改增删无效。

The Test report is invalid if being supplemented, deleted or altered.

4. 未经本单位书面同意，不得部分复制（全部复制除外）本检测报告。

Without prior written permission, the report cannot be reproduced, except in full.

5. 除非另有说明，本报告检验结果仅对来样负责。

Unless otherwise stated, the results shown in this test report refer only to the sample(s) submitted.

6. 对检测报告有异议的，应于收到报告之日起十五日内提出，逾期不予受理。

Any dispute of the report must be raised to the testing body within 15 days after the report is received, exceeding which the dispute will not be accepted.

7. 对送检样品，样品信息由委托方提供，本单位不对其真实性负责。

For the tested sample(s) submitted by the applicant, the sample information in the test report is provided by the applicant and the laboratory is not responsible for its authenticity.

8. 本检测报告仅供委托方参考，不具有对社会证明作用。

This test report is for reference only to the applicant and does not have a proof of effect for others.

